Aposición Incompleta

Falta de contacto entre por lo menos 1 strut del stent y la íntima de la pared arterial con evidencia de flujo por detrás del stent en ausencia de bifurcación.

Se ha asociado con un mayor riesgo de eventos tromboembólicos periprocedimiento.

Es un fenómeno temporal persistente , que se resuelve espontáneamente sólo en una pequeña minoría de casos y parece ser un factor de riesgo de eventos isquémicos tardíos.

Aunque aún se necesita seguimiento, estos resultados sugieren que se podría justificar una mayor duración de la terapia antiplaquetaria y un mayor seguimiento clínico [1)]

1)
Heller R, Calnan DR, Lanfranchi M, Madan N, Malek AM. Incomplete stent apposition in Enterprise stent-mediated coiling of aneurysms: persistence over time and risk of delayed ischemic events. J Neurosurg. 2013 May;118(5):1014-22. doi: 10.3171/2013.2.JNS121427. Epub 2013 Mar 15. PubMed PMID: 23495874.

From:
http://www.neurocirugiacontemporanea.com/ - **Neurocirugía Contemporánea ISSN 1988-2661**

Permanent link:
http://www.neurocirugiacontemporanea.com/doku.php?id=aposicion_incompleta

Last update: **2019/09/26 22:18**